



# जिला स्वास्थ्य समिति, औरंगाबाद

स्वास्थ्य विभाग, बिहार सरकार



## उपकरण क्रय संबंधी अल्पकालीन निविदा सूचना

राज्य स्वास्थ्य समिति, बिहार पटना द्वारा वित्तीय वर्ष 2016-17 में प्राप्त आवंटन के अन्तर्गत सदर अस्पताल, औरंगाबाद हेतु उपकरण के क्रय हेतु निर्माता कम्पनी या उनके प्राधिकृत एजेन्सी से निविदा आमंत्रित की जाती है। निर्माता कम्पनी या उनके प्राधिकृत एजेन्सी द्वारा दिनांक 25/03/2017 को अपराह्न 5:00 बजे तक निबंधित डाक/स्पीड पोस्ट द्वारा जिला स्वास्थ्य समिति, औरंगाबाद कार्यालय में निविदा जमा किया जा सकता है। निर्धारित तिथि एवं समय के बाद कोई भी निविदा स्वीकार नहीं किया जायेगा। निविदा निर्धारित नियम एवं शर्तों के अनुसार समर्पित करना होगा। प्राप्त निविदा को दिनांक 27/03/2017 को पूर्वाह्न 11:00 बजे क्रय समिति के समक्ष खोला जायेगा। निविदा खोलने की तिथि को निविदादाता स्वयं या उनके द्वारा प्राधिकृत को उपस्थित रहना अनिवार्य होगा। निविदा पूर्णतः या अंशतः निरस्त करने का अधिकार जिला स्वास्थ्य समिति, औरंगाबाद को सुरक्षित रहेगा।

निविदा की नियम एवं शर्तें तथा उपकरणों का सूची/विशिष्टता औरंगाबाद के वेबसाइट <http://aurangabad.bih.nic.in> से डाउनलोड किया जा सकता है। उपकरणों की सूची/विशिष्टता के आधार पर ही दर अंकित करना अनिवार्य होगा।

### निविदा की न्यूनतम मानक अर्हता (तकनीकी भाग)

1. प्रतिष्ठान/एजेन्सी को वाणिज्यिक निबंधन प्रमाण पत्र, सर्विस टैक्स की छाया प्रति संलग्न करना अनिवार्य होगा।
2. तकनीकी निविदा के साथ जिला स्वास्थ्य समिति, औरंगाबाद को भुगतये रूपये राशि ₹50,000/- (पचास हजार रूपये) मात्र का बैंक ड्राफ्ट संलग्न करना अनिवार्य होगा।
3. प्रतिष्ठान/एजेन्सी को अपने प्रोपराइटर/फर्म का पैन संख्या की छायाप्रति संलग्न करना होगा।
4. निविदादाता को गत तीन वर्षों का प्रोपराइटर/फर्म का इनकम टैक्स रिटर्न की छायाप्रति संलग्न करना होगा।
5. निविदादाता दुकान एवं प्रतिष्ठान अधिनियम के अन्तर्गत पंजीकृत होने का प्रमाणपत्र संलग्न करना होगा।
6. निविदादाता को गत तीन वर्षों का अर्थात् 31.03.16 तक का अंकेक्षण प्रतिवेदन का किसी Chartered Accountant के द्वारा अभिप्रमाणित छायाप्रति संलग्न करना होगा।
7. निविदादाता का विगत तीन वित्तीय वर्षों का औसतन टर्न ओवर कम से कम एक करोड़ होना चाहिए।
8. निविदादाता को सरकारी संस्थानों में Last Three Years कार्य करने/सामग्री आपूर्ति से संबंधित अनुभव होना चाहिए एवं इस हेतु कार्यदेश की छायाप्रति संलग्न करना होगा।
9. निविदादाता को Non Conviction (एवं किसी सरकारी एवं गैर सरकारी संस्था द्वारा काली सूची में दर्ज नहीं किये जाने के आशय का शपथ पत्र संलग्न करना अनिवार्य होगा।

वित्तीय भाग : निविदादाता को निविदा के वित्तीय भाग में निविदा के साथ संलग्न उपकरण के अनुसार नीचे अंकित प्रपत्र में प्रति इकाई सभी शुल्क सहित दर अंकित करना होगा। (उपस्कर/उपकरण का नाम/विशिष्टता जिला स्वास्थ्य समिति, औरंगाबाद द्वारा उपलब्ध कराये गये सूची के अनुसार ही दिया जाना अनिवार्य होगा, अन्यथा मान्य नहीं होगा।)

क्र० सं०	उपस्कर/उपकरण का नाम	उपकरणों की विशिष्टता	दर प्रति इकाई (सभी कर सहित)

सिविल सर्जन-सह-सदस्य सचिव  
जिला स्वास्थ्य समिति, औरंगाबाद

जिला पदाधिकारी-सह-अध्यक्ष,  
जिला स्वास्थ्य समिति, औरंगाबाद

## LIST OF EQUIPMENT



Sl. no.	Name of the Equipment	Specification	Rate per unit Including All Taxes
1	Anesthetic - laryngoscope magills	Should have 4 blades from ISO 13485:2013 or 2016 certified company	
2	Applanation Tonometer	Applanation type/ Based on Goldmann Tonometry principle/Should have a measuring range from 0 to 78 mmHg in steps of 2 mmHg/ Have an accuracy of $\pm 0.5$ mmHg/To be supplied with calibration Bar, Prism and tonometer mount base to fix with optics/compatible with all models of slit lamps/ spare prism/ Should be visible and clearly defined/ Labels and markings should be clear and visible.	
3	A-Scan Biometer	ISO 13485:2013 or 2016 certified company	
4	Audiometer	ISO 13485:2013 or 2016 certified company	
5	Auto Clave HP Horizontal	Horizontal Rectangular/ High Pressure/ High Vacuum steam Sterilizer/Chamber minimum of 2 feet width x 2 feet height x 4 feet length/Should be operated on 400-440V, 3 phase with neutral, AC power supply/ Manufactured as per IS specifications Mark ISI:3829 or BIS standard	
6	Auto Clave HP Vertical (2 bin)	Vertical 2 bin capacity/ High Pressure/ High Vacuum steam SterilizerShould be operated on 400-440V, 3 phase with neutral, AC power supply/ Manufactured as per IS specifications Mark ISI:3829 or BIS standard.	
7	Auto Refractometer	Refractive measurement sphere from -25 to +22D in steps of 0.12 / 0.25 D/Refractive measurement cylinder from -10 to +10 in steps of 0.12 /0.25 D./Refractive measurement axis angle from 1 to 180° in steps of 1°/At least 0, 12 and 13.5 vertex distance/ Should measure a minimum pupil diameter of 2.5mm/ Should have maximum pupil distance measurement of 85 mm in steps of 1mm/ Should have at least 5 inches colour LCD display/ Should have vertically adjustable chin rest of at least $\pm 25$ mm/ Should have an inbuilt thermal printer/ Should be supplied with motorized table/ Should provide printing paper – 10 Nos. and Chin rest paper – 200 Nos/ Should operate on 200 to 240 Vac, 50Hz input supply.	
8	Autoclave vertical single bin	Vertical single bin capacity/ High Pressure/ High Vacuum steam SterilizerShould be operated on 400-440V, 3 phase with neutral, AC power supply/ Manufactured as per IS specifications Mark ISI:3829 or BIS standard.	
9	B.P.apparatus stand model	Portable mercurial type/ have ISI mark/ Measuring range from 0 to 300 mmHg.	
10	Binocular Microscope	USFDA or European CE approved product/ Power input to be 220-240VAC, 50Hz fitted with Indian plug.	
11	Binomags	Binocular and simple magnification.	
12	Bowl Sterilizer	Electrical big size.	
13	Bowl Sterilizer Medium	Electrical medium size.	

*Patna  
15/08/24*

*MO Suresh Kumar  
11.08.24*

*11/08/24  
JS*

*3/11/24*

